

CUSTOMIZED GUIDE FOR FEMORAL COMPONENT POSITIONING IN HIP RESURFACING ARTHROPLASTY

GUIA PERSONALIZADA PARA POSICIONAMENTO DE COMPONENTE FEMORAL PARA *RESURFACING* EM ARTROPLASTIA DO QUADRIL

CHATURONG PORNRATTANAMANEWONG¹, RAPEEPAT NARKBUNNAM¹, KEERATI CHAREANCHOLVANICH¹

1. Mahidol University, Faculty of Medicine, Siriraj Hospital, Department of Orthopedic Surgery, Bangkok, Thailand.

ABSTRACT

Objective: To prove the accuracy of a customized guide developed according to our method. **Methods:** This customized guide was developed from a three-dimensional model of proximal femur reconstructed using computed tomography data. Based on the new technique, the position of the guide pin insertion was selected and adjusted using the reference of the anatomical femoral neck axis. The customized guide consists of a hemispheric covering designed to fit the posterior part of the femoral neck. The performance of the customized guide was tested in eight patients scheduled for total hip arthroplasty. The stability of the customized guide was assessed by orthopedic surgeons. An intraoperative image intensifier was used to assess the accuracy. **Results:** The customized guide was stabilized with full contact and was fixed in place in all patients. The mean angular deviations in relation to the what was planned in anteroposterior and lateral hip radiographs were $0.5^\circ \pm 1.8^\circ$ in valgus and $1.0^\circ \pm 2.4^\circ$ in retroversion, respectively. **Conclusion:** From this pilot test, the authors suggest that the proposed technique could be applied as a customized guide to the positioning device for hip resurfacing arthroplasty with acceptable accuracy and user-friendly interface. **Level of Evidence IV, Cases Series.**

Keywords: Arthroplasty, replacement, hip/instrumentation. Femur neck. Imaging, three-dimensional. Surgery, computer-assisted. Prosthesis design.

RESUMO

Objetivo: Comprovar a precisão de uma guia personalizada desenvolvida a partir de nosso método. **Métodos:** Esta guia personalizada foi desenvolvida a partir do modelo tridimensional da parte proximal do fêmur reconstruída usando dados de tomografia computadorizada. Com base na nova técnica, a posição de inserção do pino da guia foi selecionada e ajustada usando a referência do eixo anatômico do colo do fêmur. A guia personalizada consiste em um revestimento hemisférico projetado para encaixar na parte posterior do colo do fêmur. O desempenho da guia personalizada foi testado em oito pacientes que seriam submetidos à artroplastia total do quadril. A estabilidade da guia personalizada foi avaliada por cirurgiões ortopedistas. Para avaliar a precisão, usou-se um intensificador de imagem intraoperatório. **Resultados:** A guia personalizada foi estabilizada com contato total e foi fixada em todos os pacientes. Os desvios angulares médios com relação ao planejado nas radiografias anteroposteriores e laterais do quadril foram de $0,5^\circ \pm 1,8^\circ$ em valgo e $1,0^\circ \pm 2,4^\circ$ em retroversão, respectivamente. **Conclusão:** A partir deste teste piloto, os autores sugerem que a técnica proposta poderia ser aplicada como guia personalizada para o dispositivo de posicionamento para resurfacing em artroplastia de quadril com aceitável precisão e interface amigável. **Nível de Evidência IV, Série de Casos.**

Descritores: Artroplastia de quadril/instrumentação. Colo do fêmur. Imagem tridimensional. Cirurgia assistida por computador. Desenho de prótese.

Citation: Pornrattanamaneewong C, Narkbunnam R, Chareancholvanich K. Customized guide for femoral component positioning in hip resurfacing arthroplasty. *Acta Ortop Bras.* [online]. 2017;25(2):103-6. Available from URL: <http://www.scielo.br/aob>.

INTRODUCTION

Hip resurfacing arthroplasty (HRA) is an alternative to total hip arthroplasty (THA). The advantages of this procedure include preservation of the femoral bone stock,¹ minimized dislocation rate,² and improved range of motion.³ However, HRA is a technically demanding procedure and femoral neck fracture has been documented as the most common cause of early failure.⁴ This complication is related to varus malposition of the femoral component and superior notching of femoral neck.^{5,6} Accurate positioning of the femoral component has been reported in association with successful long-term outcomes.⁷ Optimal alignment

traditionally is achieved using manual devices, and accuracy relies largely on visual inspection and the surgeon's experience. Computer assisted navigation can increase the accuracy of femoral guide pin insertion compared to conventional instrumentation.^{8,9} Nevertheless, it has distinctive disadvantages, including increased surgical time and cost.⁸ Patient-specific instrumentation for HRA is a novel device fabricated using rapid prototyping technology (RP). Computed tomography (CT) scanning provides individual 3D geometric anatomy data to construct a patient-specific instrument. The instrument is used to guide the position of pin insertion to avoid malpositioning of the femoral component. The most important reference axis for

All the authors declare that there is no potential conflict of interest referring to this article.

Study conducted at Siriraj Hospital, Bangkok, Thailand

Correspondence: Keerati Chareancholvanich. Siriraj Road, Bangkoknoi, Bangkoknoi, Bangkok, Thailand 10700. Department of Orthopedic Surgery, Faculty of Medicine, Siriraj Hospital, Mahidol University, Bangkok, Thailand. keesi93@gmail.com

Article received in 08/02/2016, approved in 09/09/2016.

Acta Ortop Bras. 2017;25(2):103-6

determining guide pin direction is the femoral neck axis (FNA). In clinical practice, determining the true FNA can still be problematic. Although several patient-specific guides (PSG) have been proposed in the literature and demonstrated utility with good accuracy,¹⁰⁻¹⁴ few studies state the method for defining true FNA.^{11,14} To our knowledge, the best-known technique was developed by Mahaisavariya et al.,¹⁵ who established a method for geometrical assessment of the proximal femur in three dimensions. This method uses CT images combined with reverse engineering to obtain the 3D geometry of the proximal femur. This technique can be applied to HRA in order to identify the true FNA. Consequently, the objectives of this study were to use this technique to design a PSG to assist in femoral component positioning in HRA, as well as assess the accuracy of this device.

PATIENTS AND METHODS

Between May and August 2011 we recruited eight patients scheduled for unilateral primary THA to participate in the study. Preoperative CT scans of the hip were performed with a 1-mm slice thickness. The axial cross-sectional images of the body were formatted into Digital Imaging and Communications in Medicine (DICOM) files and transferred to National Metal and Materials Technology Center (MTEC, National Science and Technology Development Agency, Pathumthani, Thailand). Medical imaging processing software (Mimics, Materialise N.V., Belgium) was used to convert the set of DICOM files to a three-dimensional image of hip. In the reconstruction process, the stack of DICOM files was sequenced in such the way that the relative proximal cross-section images were above the distal cross-sectional images. This allowed the femoral head to be oriented proximally to the femoral shaft. Each image was within the threshold range for Hounsfield unit (HU) values to capture bone density. Images of the proximal femoral region were separated from other bones using a region-by-region growing algorithm.¹⁵ The captured boundaries of the proximal femoral region were interpolated in a 3D computer aided design (CAD) model of the proximal femur, as illustrated in Figure 1. PSG design can achieve success in HRA because it can precisely determine the true FNA. Nevertheless, if the femoral head was severely deformed, the mirror image technique from the contralateral side was used to calculate this axis. The engineers at MTEC developed a technique to derive true FNA as follows: the least square regression of ellipse and sphere was performed at the femoral neck region and femoral head, respectively. The centers of the ellipse and sphere together derived the line using the linear regression technique; this derived line was the true FNA. However, since is difficult to determine the correct cross-section plane at the femoral neck used for least square regression of the ellipse, the iteration of the aforementioned least square regression technique was performed until the FNA resulting from the current iteration was no more than $\pm 0.5^\circ$ different from the FNA resulting from the previous iteration.¹⁵




Figure 1. Three-dimensional graphic model of the proximal femur.

The guide pin was planned following the true FNA for 5 cases and plus additional overcorrection of 5 degrees valgus for 3 cases. (Figure 2) The PSG was hemispherical to cover the femoral head, and was designed to be placed on the posterior part of the femoral head and neck. (Figure 3) In addition, the PSG contained the sleeve to control drilling direction, allowing a 3.2-mm diameter guide pin to be inserted through it. The primary goal of these devices was to securely fit the femoral neck, because of concerns related to the indistinct contours of articular cartilage in the CT images. The PSG was fabricated using a RP machine developed in our facility using an acrylate resin biocompatible for bone contact. The PSG was then polished and cleaned to remove residual particles, and sterilized using gamma radiation. Before PSG was used in surgery, trials were performed in all cases. The PSG were tested to assess contact with the physical model of the proximal femur, stability of fixture, and contact clearance, as well as evaluate the direction of the guide after drilling. These assessments were made by the orthopedic surgeons as well as the engineers. This study was approved by the institutional review board of Siriraj Hospital. All preoperative CT scans were done within 4 weeks prior to surgery. All surgical procedures were performed by the senior author (CK). The patient was placed in a lateral decubitus position, and a posterolateral approach of the hip was performed. After dislocation of the femoral head, the PSG was wrapped around the posterior part of femoral head and neck and locked in a stable snap-fit position. (Figure 4) The contact obtained at the neck portion and the stability of fixture was graded by the surgeons (full contact/unmovable, partial contact/unmovable, and partial contact/movable). A 3.2-mm-diameter guide pin was inserted via the pinhole and passed through the femoral neck. After removing the PSG, another end of the pin was cut at the level of 2 mm above the femoral head. The femoral head containing the pin was relocated. Pin alignment was assessed using




Figure 2. Adjusted axis (AA) to the direction of 5° valgus from the true femoral neck axis (FNA).




Figure 3. Design of patient-specific guide, (a) posterior and (b) superior view of proximal femur.




Figure 4. Intraoperative application of the patient-specific guide.

the intraoperative image intensifier. For the anteroposterior (AP) view, the image intensifier was positioned perpendicular to the hip with femoral internal rotation of 15°. Without moving the image intensifier, the lateral view was obtained with a hip position of 45° flexion, 45° abduction, and 30° external rotation. After radiologic examination, the femoral head was dislocated again and the guide pin was removed. The remainder of the THA procedure was carried out as usual. No intra- or postoperative complications occurred in this series.

Radiographic evaluation

Two blinded assessors were assigned to evaluate the radiographs. From the AP view, we modified Muller's method¹⁶ to determine FNA as follows: the center of the femoral head was located with a circle. Reference points for the circle arc were the inferomedial and inferolateral border. The point of deepest concavity on the lateral border of femoral neck was marked. Another circle arc using the center of femoral head as the center was drawn. The points where the circle intersected the femoral neck were connected and defined as the transcervical line. Another line drawn perpendicular to the transcervical line through the center of the femoral head represented the FNA. This method was also applied in the lateral view, but used the anteroinferior and posteroinferior border to define the femoral head. (Figure 5) The direction of the guide pin was compared to the FNA in either the AP or lateral radiographs. Deviations between these two lines were defined as angular deviations; varus, neutral or valgus angulation in the AP view and anteversion, neutral or retroversion in the lateral view.




Figure 5. Radiographic assessment of guide pin position in anteroposterior (left side) and lateral (right side) radiographs of the hip; the center of the femoral head (C) is located within circle A. Another circle arc (B) was drawn using the center of the femoral head (C) as the center. Points a and b, where circle B intersects the deepest concave point of the femoral neck, are connected and defined as the transcervical line (ab line). The line drawn perpendicular to the transcervical line through the center of the femoral head (C) is represented as the femoral neck axis (FNA).

Statistical analysis

Statistical analysis in this study was performed using SPSS version 18.0 software (SPSS Inc., Chicago, IL). Mean \pm standard deviation (SD) was used to explain the descriptive statistics. Intraclass correlation coefficients (ICC) were used to assess the intra- and inter-rater reliability of radiographic measurements.

RESULTS

The characteristics of the patients and details of the assessed outcome are presented in Table 1. Mean patient age was 47.0 ± 12.1 years. The majority of the patients were diagnosed with osteonecrosis. The PSG was stabilized with full contact and was unmovable in all patients. The mean angular deviations from planning in the AP and lateral radiographs were 0.5° valgus $\pm 1.8^\circ$ and 1.0° retroversion $\pm 2.4^\circ$ directions, respectively. The ICC for inter-rater reliability was 0.83 and 0.91, while ICC for intra-rater reliability was 0.91 and 0.96 for angular deviation assessment in AP and lateral radiographs, respectively.

Table 1. Patient characteristics and radiographic outcomes.

Number	Sex	Age (yrs)	Side	Diagnosis	Anteroposterior radiograph		Lateral radiograph	
					Aim	Angular deviation	Aim	Angular deviation
1	Female	26	Right	ON	0°	3° varus	0°	3° anteversion
2	Female	52	Left	ON	0°	1° valgus	0°	2° retroversion
3	Female	38	Right	PVNS	0°	neutral	0°	4° retroversion
4	Female	62	Right	OA	0°	neutral	0°	3° retroversion
5	Female	54	Left	OA	0°	3° valgus	0°	neutral
6	Female	58	Left	DDH	5° valgus	6° valgus	0°	3° retroversion
7	Female	48	Right	ON	5° valgus	7° valgus	0°	neutral
8	Male	38	Right	ON	5° valgus	5° valgus	0°	1° anteversion

ON = osteonecrosis, PVNS = pigmented villonodular synovitis. OA = osteoarthritis, DDH = developmental dysplasia of the hip.

DISCUSSION

Various computer aided design and manufacturing technologies (CAD/CAM) were employed in this study to obtain the PSG used for HRA surgery. Proper selection of HU values in image processing along with reverse engineering technologies permit an accurate geometric model of the proximal femur to be constructed based on CT images. Determination of the true FNA based on the 3D proximal femur involves geometric approximation using various least square regressions, i.e. fit ellipse, fit sphere, and fit line. Until recently, the technologies described in this paper were not available for HRA, and FNA for most procedures was determined by the surgeon's skill, with varying results. These results may vary because the anthropometry of the Thai proximal femur, especially FNA, presents a wide range of values (110°–140°).¹⁵ Therefore, the specific instrument for HRA presented in this study was reasonable. The PSG we developed is meant to facilitate the surgical procedure by precisely determining the true FNA and reducing trial and the use of radiography. Unlike conventional radiographic imagery, three-dimensional CAD allows the true FNA to be determined. In order to demonstrate the accuracy of FNA determined through the PSG, we assessed the position during surgery and the angles measured in the AP and lateral views. The results show that the PSG has an acceptable coronal alignment accuracy of $\pm 5^\circ$.¹⁷ In the sagittal alignment, although there was no evidence of acceptable alignment for anteversion, little angular deviation was shown from the study. True FNA is important for placing the femoral component in HRA. Most studies investigating PSG do not mention how to determine

FNA^{10,12,13,17,18} only one method, the so-called translucent cylinder method, appears in the previous literature. A translucent cylinder is created and superimposed on the 3D femoral model, and the position, orientation, and size of this cylinder are adjusted to identify FNA and neck diameter.^{11,14} Using the translucent cylinder method, Kunz et al.¹¹ reported an angular deviation of 1.14° in coronal alignment and 4.49° in sagittal alignment. Du et al.¹⁴ assessed reports of angular deviation after using PSG with translucent cylinder method were displayed, and the results concentrated on the stem-shaft angle (SSA) difference between PSG and conventional instruments. Different PSG designs and surgical approaches have been proposed in the literature. (Table 2) Most authors used the PSG via the posterior approach to the hip. Acceptable coronal alignment was demonstrated in most studies,^{10-14,18} except for Olsen et al.,¹⁷ who reported an angular deviation of 6.4 ± 2.9° in the coronal plane. In the sagittal plane, a maximum angular deviation of 4.49° was reported using 3D CT navigation assessment.¹¹ In the current study, the authors developed the PSG according to the true FNA obtained from CAD in conjunction with various regression techniques. Compared to other studies, the PSG we developed can be applied with good stability and provides acceptable accuracy

in guide pin placement for HRA. (Table 2) Nevertheless, there are several limitations in this study; first, this study is preliminary and was only conducted in a small group of THA patients. Future studies should investigate the use of this device in HRA in larger groups. Second, PSG was designed only for the posterior approach. During surgery, damage to the vessels in the posterior capsule of the hip may cause avascular necrosis of the femoral neck, resulting in femoral neck fracture.¹⁹ Finally, the authors assessed the alignment using an image intensifier. Although we try to control the position of hip and leg, some imaging error may occur. Intraoperative CT scanning is the best option in this situation, but is not available in our institute.

CONCLUSION

This study presents the use of CAD/CAM in conjunction with various least square regression techniques to determine true FNA and develop and fabricate a PSG for femoral component positioning in HRA. The initial results from eight patients using CT based PSG are encouraging. The shape of the PSG was applied to the femoral neck and a secure fit was obtained, and accurate guide pin insertion using this device was verified.

Table 2. Comparison with previous studies investing the accuracy of patient-specific guides.

Study	Published year	Subjects	Surgical approach of the hip	Outcome measurement	Angular deviation (mean SD)	
					Coronal plane	Sagittal plane
Kunz et al. ¹¹	2010	45 HRA	Anterolateral	CT navigation	1.14°	4.49°
Raaijmakers et al. ¹²	2010	5 THA	Anterolateral	Optical scan	Maximal 2.9°	
Zhang et al. ¹³	2011	10 HRA	Posterior	Image intensifier	1.3 ± 1.0°	NA
Audenaert et al. ¹⁰	2011	5 cadavers	Posterior	CT	4.05 ± 1.84°	
Du et al. ¹⁴	2013	16 HRA	Posterior	Plain radiographs	NA	
Kitada et al. ¹⁸	2013	12 synthetic femoral models	Posterior	CT	2.5 ± 2.4°	1.5 ± 2.3°
Olsen et al. ¹⁷	2009	6 cadavers	Direct lateral Posterior	Plain radiographs	6.4 ± 2.9°	1.0 ± 0.4°
Current study		8 THA	Posterior	Image intensifier	0.5 ± 1.8°	1.0 ± 2.4°

SD = standard deviation, HRA = hip resurfacing arthroplasty, THA = total hip arthroplasty, CT = computed tomography, NA = not applicable.

AUTHORS' CONTRIBUTIONS: Each author contributed individually and significantly to the development of the manuscript. CP (0000-0001-8376-014)* and KC (0000-0003-2837-9182)* were the main contributors in drafting the manuscript. KC performed surgery, followed patients, and gathered clinical data. CP and RN (0000-0002-3564-1700)* evaluated the data from the statistical analysis. CP, RN, and KC performed the literature search, reviewed the manuscript, and contributed to the intellectual concept of the study. *ORCID (Open Researcher and Contributor ID).

REFERENCES

- Crawford JR, Palmer SJ, Wimhurst JA, Villar RN. Bone loss at hip resurfacing: a comparison with total hip arthroplasty. *Hip Int.* 2005;15(4):195-8.
- McMinn D, Treacy R, Lin K, Pynsent P. Metal on metal surface replacement of the hip. Experience of the McMinn prosthesis. *Clin Orthop Relat Res.* 1996;(329Suppl):S89-98.
- Chandler DR, Glousman R, Hull D, McGuire PJ, Kim IS, Clarke IC, et al. Prosthetic hip range of motion and impingement. The effects of head and neck geometry. *Clin Orthop Relat Res.* 1982;(166):284-91.
- Macpherson GJ, Breusch SJ. Metal-on-metal hip resurfacing: a critical review. *Arch Orthop Trauma Surg.* 2011;131(1):101-10.
- Davis ET, Olsen M, Zdero R, Waddell JP, Schemitsch EH. Femoral neck fracture following hip resurfacing: the effect of alignment of the femoral component. *J Bone Joint Surg Br.* 2008;90(11):1522-7.
- Vail TP, Glisson RR, Dominguez DE, Kitaoka K, Ottaviano D. Position of hip resurfacing component affects strain and resistance to fracture in the femoral neck. *J Bone Joint Surg Am.* 2008;90(9):1951-60.
- Beaulé PE, Lee JL, Le Duff MJ, Amstutz HC, Ebrahmdadeh E. Orientation of the femoral component in surface arthroplasty of the hip. A biomechanical and clinical analysis. *J Bone Joint Surg Am.* 2004;86(9):2015-21.
- Resubal JR, Morgan DA. Computer-assisted vs conventional mechanical jig technique in hip resurfacing arthroplasty. *J Arthroplasty.* 2009;24(3):341-50.
- Schnurr C, Michael JW, Eysel P, König DP. Imageless navigation of hip resurfacing arthroplasty increases the implant accuracy. *Int Orthop.* 2009;33(2):365-72.
- Audenaert E, De Smedt K, Gelaude F, Clijmans T, Pattyn C, Geebelen B. A custom-made guide for femoral component positioning in hip resurfacing arthroplasty: development and validation study. *Comput Aided Surg.* 2011;16(6):304-9.
- Kunz M, Rudan JF, Xenoyannis GL, Ellis RE. Computer-assisted hip resurfacing using individualized drill templates. *J Arthroplasty.* 2010;25(4):600-6.
- Raaijmakers M, Gelaude F, De Smedt K, Clijmans T, Dille J, Mulier M. A custom-made guide-wire positioning device for hip surface replacement arthroplasty: description and first results. *BMC Musculoskelet Disord.* 2010;11:161.
- Zhang YZ, Lu S, Yang Y, Xu YQ, Li YB, Pei GX. Design and primary application of computer-assisted, patient-specific navigational templates in metal-on-metal hip resurfacing arthroplasty. *J Arthroplasty.* 2011;26(7):1083-7.
- Du H, Tian XX, Li TS, Yang JS, Li KH, Pei GX, et al. Use of patient-specific templates in hip resurfacing arthroplasty: experience from sixteen cases. *Int Orthop.* 2013;37(5):777-82.
- Mahaisavariya B, Sitthiseripratip K, Tongdee T, Bohez EL, Vander Sloten J, et al. Morphological study of the proximal femur: a new method of geometrical assessment using 3-dimensional reverse engineering. *Med Eng Phys.* 2002;24(9):617-22.
- Muller ME. *Die huftnahen Femurosteotomien.* Stuttgart: Georg Thieme Verlag; 1971.
- Olsen M, Davis ET, Waddell JP, Schemitsch EH. Imageless computer navigation for placement of the femoral component in resurfacing arthroplasty of the hip. *J Bone Joint Surg Br.* 2009;91(3):310-5.
- Kitada M, Sakai T, Murase T, Hanada T, Nakamura N, Sugano N. Validation of the femoral component placement during hip resurfacing: a comparison between the conventional jig, patient-specific template, and CT-based navigation. *Int J Med Robot.* 2013;9(2):223-9.
- Nork SE, Schär M, Pfander G, Beck M, Djonov V, Ganz R, et al. Anatomic considerations for the choice of surgical approach for hip resurfacing arthroplasty. *Orthop Clin North Am.* 2005;36(2):163-70.